



केन्द्रीय रेशम प्रौद्योगिक अनुसंधान संस्थान

केन्द्रीय रेशम बोर्ड

(यस्त्र मंत्रालय- भारत सरकार)

बी.टी.एम. लेआउट मडिवाला बेंगलूरु- 560068


CENTRAL SILK TECHNOLOGICAL RESEARCH INSTITUTE

CENTRAL SILK BOARD

(Ministry of Textiles - Govt. of India)

B.T.M. Layout, Madivala, Bengaluru-560068

Phone: 080-26685238, 26282190, Fax: 080-26680435, Email: [cstriban.csb@nic.in](mailto:cstriban.csb@nic.in), web: [www.cstri.res.in](http://www.cstri.res.in)



No. CSB/CSTRI/2(3)/2024-Stores/Guwahati/PS

Date: 09.09.2025

To,

Sir,

Sub: Inviting Quotations through e-procurement portal for procurement of Multi-Purpose Pressure Sensors for Stockings for RSTRS, CSB-CSTRI, Guwahati - Reg.

\*\*\*\*\*

With reference to the subject cited above, I am to inform that this Institute invites quotation in two bid system through e-procurement portal <https://eprocure.gov.in> for supply of machinery as mentioned below:-

#	NAME OF THE EQUIPMENT	Specification	QTY.	DESTINATION
1.	Multi-Purpose Pressure Sensors for Stockings	Detailed Specification is enclosed	1 No.	The Scientist-D & In-charge, Regional Silk Technological Research Station, Central Silk Board, Khanapara, Muga Farm, Guwahati, Dist : Kamrup, Assam-781022.

**TERMS & CONDITION :**

1. The quotation in **two bid (Technical Bid & Financial Bid)** system should be uploaded in e-procurement only (<https://eprocure.gov.in>). The last date for uploading the bid is **07.10.2025** upto **1.00 P.M.** and the bid will be opened on **08.10.2025** at **1.00 P.M.**
2. The rates quoted should be valid for 180 days. The equipment should be supplied within **30** days from the date of purchase order. No deviation is permitted in this aspect.

3. The equipment should be supplied to "The Scientist-D & In-charge, Regional Silk Technological Research Station, Central Silk Board, Khanapara, Muga Farm, Guwahati, Dist : Kamrup, Assam- 781022".

4. In the Technical Bid:

- a. The other important documents sought in the e-procurement tender should be uploaded online.
  - b. **For Non-MSME bidders/firms:** Fixed EMD of Rs. 3,000/- in the form of Demand Draft drawn in favour of the Director, CSTRI, Bangalore should reach this office by post / in person before the opening of technical bid. i.e on or before **08.10.2025** at 1.00 P.M.. The EMD should be sent in a sealed cover superscribing "EMD for Supply of **Multi-Purpose Pressure Sensors for Stockings**". A scanned copy of the Demand Draft should be uploaded along with technical bid. The EMD does not carry any interest. Please note that cheques will not be accepted. Quotation received without EMD will summarily be rejected.
  - c. **For MSME bidders/firms:** EMD amount is exempted for MSME units. In such cases, scanned copy of **MSME certificate** has to be uploaded along with technical bid. Further, such MSME bidders should also submit the duly filled **Bid Securing Declaration form** alongwith the technical bid (as per the format enclosed) in lieu of EMD.
  - d. The bidders should submit the Bank Details i.e. Bank A/c.No., Bank Name, Branch, IFSC code. Mobile No. E-Mail ID, Aadhaar, etc. for releasing the EMD and effecting the payment.
  - e. The bidders should submit an undertaking for not being blacklisted in the letter head.
  - f. The supplier should submit the experience certificate/purchase orders relating to the supply of **items to government/private institutions/organizations**.
  - g. The bidders should submit an undertaking that the above equipment to be supplied shall match with the specifications as attached.
  - h. The Bidder should submit the copy of GST Certificate, ITR Returns, PAN number of the firm, an affidavit regarding the correctness of bid, Aadhaar along with the technical bid.
5. The Supplier shall provide Warranty for a minimum period of 12 months from the date of installation of the equipment / machineries. Any defects noticed in the equipment /machineries which are unintentional shall be set right by the supplier at free of cost for good performance of the equipments.
6. The equipment should be supplied strictly in conformity of the specifications as attached, failing which, the equipment has to be taken back at your own cost.

7. For delay & non supply of the item within the stipulated time, liquidated damages @ 0.5% per week will be deducted from the bill subject to a maximum of 10% of the total cost of the equipment.
8. The performance security equivalent to 3% of the total contract value in the form of Bank Guarantee issued by any Nationalized Bank which shall be valid till three months after the expiry of the warranty period should be furnished to this Institute within 10 days from the date of receipt of purchase order. In case of non receipt of the Bank Guarantee within the stipulated time, the said amount will be deducted while effecting the final payment.
9. No advance payment will be made. The payment will be released only after delivery of the equipment in good condition as per our specification and commissioning.
10. **In the financial bid**, the bidder has to furnish the basic rate which includes transportation charges, Commissioning, insurance, loading and unloading, etc. except GST that has to be mentioned separately in the BOQ format only. There should not be any kind of hidden charges.
11. The bid is to be quoted in Indian Rupee only.
12. The applicable GST-TDS will be deducted as per the prevailing rules.
13. **Deviation in any of the above terms and conditions\* will result in automatic cancellation of the bid.**
14. Director, CSTRI, Central Silk Board, Bangalore reserves right to accept or reject any bid without assigning any reason. The decision of the Director is final and no correspondence will be entertained from any party in this regard.

Yours faithfully,

  
[डॉ. एस. पेरियासामी / Dr. S. PERIYASAMY]

निदेशक / DIRECTOR



## BID SECURING DECLARATION FORM

Tender No. **CSB/CSTRI/2(3)/2024-Stores/Guwahati/PS**

Date:

To  
Director,  
CSTRI,  
BANGALORE

I/We, The undersigned, declare that :

I/We understand that, according to your conditions, bids must be supported by a Bid Securing Declaration.

I/We accept that I/We may be disqualified from bidding for any contract with you for a period of one year from the date of notification if I am/We are in a breach of any obligation under the bid conditions, because I / We

- a) have withdrawn / modified/amended, impairs or derogates from the tender, my / our Bid during the period of bid validity specified in the form of Bid ; or
- b) having been notified of the acceptance of our Bid by the purchaser during the period of bid validity (i) fail or reuse to execute the contract, if required, or (ii) fail or reuse to furnish the Performance Security, in accordance with the Instructions to Bidders.

I/We understand this Bid securing Declaration shall cease to be valid if I am / We are not the successful Bidder, upon the earlier of (i) the receipt of your notification of the name of the successful Bidder; or (ii) thirty days after the expiration of the validity of my / our Bid.

Signed:

(Insert Signature of person whose name and capacity are shown)

In the capacity of

(Insert legal capacity of person signing the Bid Securing Declaration)

Name:

(Insert complete name of person signing the Bid Securing Declaration)

Duly authorized to sign the bid for on behalf of (Insert complete name of Bidder)

Dated on \_\_\_\_\_ day of \_\_\_\_\_ (insert date of signing)

Corporate Seal (where appropriate)

(Note: In case of Joint Venture, the Bid Securing Declaration must be in the name of all partners to the Joint Venture that submits the bid)

# **Specification of Multi-Purpose Pressure Sensors for Stockings**

**Technical specification:** Suitable for measuring pressure exerted by compression bandages and elastic stockings.

To be supplied with two sensors – one large and one small.

**Small sensor suitable** for measurement of sub-bandage pressure under compression bandages and elastic stockings – including:

- Checking of sub-bandage pressure and verification of different sub-bandage pressure on and behind the malleolus.
- Checking pressure under elastic stockings.

**Large sensor suitable for**

- Checking the sitting position of a patient – with and without pressure-relieving aid.
- Checking of interface pressure at different positioning methods – e.g., 30 or 90 degrees on the hip.

**Accurate pressure measurements** ( $< \pm 10$  mmHg)

**Easy and user-friendly operation**